

SimplexTAS® 101 Reagent Cartridge

Control information

TG

1. Reagent Cartridge information

Subject	Contents
Item	Total glycerides
Lot. No.	0100B
Expired date	2021. 7. 8

2. Control information

Subject	Contents
Item	Lyphochek® Assayed Chemistry Control
REF No.	C-310-5 Level 1 C-315-5 Level 2
Lot. No.	26460
Manufacturer	Bio-rad

3. Control range

Units	Level 1 - 26461		Level 2 - 26462	
	Mean	Range	Mean	Range
mg/dL	151	129 ~ 174	80	68 ~ 92
mmol/L	1.71	1.46 ~ 1.97	0.90	0.77 ~ 1.04



TASCOM Co., Ltd

Anyang Megavalley, 268, Hagui-ro, Dongan-gu, Anyang-si, Gyeonggi-do, 14056, Korea

Email: info@tascom.org

Website: www.simplextas.com