
















# EIGHTCHECK-3WP ASSAY SHEET














## Normal Level

### For XP series

**LOT** 4111 0822 

Parameters		Mean	Limit %	Limit #	Range	Barcode
WBC	[10 <sup>3</sup> /μL]	7.0	7.0	0.5	7.5	
					6.5	
RBC	[10 <sup>6</sup> /μL]	4.47	4.0	0.18	4.65	
					4.29	
HGB	[g/dL]	12.0	4.0	0.5	12.5	
					11.5	
	[mmol/L]	7.5	4.0	0.3	7.8	
					7.2	
HCT	[%]	34.4	7.5	2.6	37.0	
					31.8	
	[L/L]	0.344	7.5	0.026	0.370	
					0.318	
MCV	[fL]	76.9	5.0	3.8	80.7	
					73.1	
MCH	[pg]	26.8	5.5	1.5	28.3	
					25.3	
	[amol]	1666	5.5	92	1758	
					1574	
MCHC	[g/dL]	34.9	7.0	2.4	37.3	
					32.5	
	[mmol/L]	21.6	7.0	1.5	23.1	
					20.1	
PLT	[10 <sup>3</sup> /μL]	226	15	34	260	
					192	
W-SCR / LYM	[%]	30.9	15.0	4.6	35.5	
					26.3	

 27-Jul-2024 

Parameters		Mean	Limit %	Limit #	Range	Barcode
W-MCR / MXD	[%]	11.5	25.0	2.9	14.4	
					8.6	
W-LCR / NEUT <sup>1)</sup>	[%]	57.6	7.0	4.0	61.6	
					53.6	
W-SCC / LYM <sup>1)</sup>	[10 <sup>3</sup> /μL]	2.2	15.0	0.3	2.5	
					1.9	
W-MCC / MXD <sup>1)</sup>	[10 <sup>3</sup> /μL]	0.8	25.0	0.2	1.0	
					0.6	
W-LCC / NEUT <sup>1)</sup>	[10 <sup>3</sup> /μL]	4.0	10.0	0.4	4.4	
					3.6	
W-SMV	[fL]	59.0	18.0	10.6	69.6	
					48.4	
W-LMV	[fL]	189.6	15.0	28.4	218.0	
					161.2	
RDW-SD	[fL]	38.0	15.0	5.7	43.7	
					32.3	
RDW-CV	[%]	12.1	15.0	1.8	13.9	
					10.3	
PDW	[fL]	8.4	10.0	0.8	9.2	
					7.6	
MPV	[fL]	9.2	8.0	0.7	9.9	
					8.5	
P-LCR	[%]	14.5	60.0	8.7	23.2	
					5.8	
PCT	[%]	0.21	24.0	0.05	0.26	
					0.16	

1) These parameters values should be used as a reference only.