
















# EIGHTCHECK-3WP ASSAY SHEET














## Normal Level

### For XP series

**LOT** 4027 0822 

Parameters		Mean	Limit %	Limit #	Range	Barcode
WBC	[10 <sup>3</sup> /μL]	7.5	7.0	0.5	8.0	
					7.0	
RBC	[10 <sup>6</sup> /μL]	4.42	4.0	0.18	4.60	
					4.24	
HGB	[g/dL]	12.4	4.0	0.5	12.9	
					11.9	
	[mmol/L]	7.7	4.0	0.3	8.0	
					7.4	
HCT	[%]	35.0	7.5	2.6	37.6	
					32.4	
	[L/L]	0.350	7.5	0.026	0.376	
					0.324	
MCV	[fL]	79.1	5.0	4.0	83.1	
					75.1	
MCH	[pg]	28.1	5.5	1.5	29.6	
					26.6	
	[amol]	1741	5.5	96	1837	
					1645	
MCHC	[g/dL]	35.4	7.0	2.5	37.9	
					32.9	
	[mmol/L]	22.0	7.0	1.5	23.5	
					20.5	
PLT	[10 <sup>3</sup> /μL]	232	15	35	267	
					197	
W-SCR / LYM	[%]	30.0	15.0	4.5	34.5	
					25.5	

 04-May-2024 

Parameters		Mean	Limit %	Limit #	Range	Barcode
W-MCR / MXD	[%]	11.7	25.0	2.9	14.6	
					8.8	
W-LCR / NEUT <sup>1)</sup>	[%]	58.3	7.0	4.1	62.4	
					54.2	
W-SCC / LYM <sup>1)</sup>	[10 <sup>3</sup> /μL]	2.3	15.0	0.3	2.6	
					2.0	
W-MCC / MXD <sup>1)</sup>	[10 <sup>3</sup> /μL]	0.9	25.0	0.2	1.1	
					0.7	
W-LCC / NEUT <sup>1)</sup>	[10 <sup>3</sup> /μL]	4.3	10.0	0.4	4.7	
					3.9	
W-SMV	[fL]	59.9	18.0	10.8	70.7	
					49.1	
W-LMV	[fL]	193.5	15.0	29.0	222.5	
					164.5	
RDW-SD	[fL]	35.6	15.0	5.3	40.9	
					30.3	
RDW-CV	[%]	10.7	15.0	1.6	12.3	
					9.1	
PDW	[fL]	9.0	10.0	0.9	9.9	
					8.1	
MPV	[fL]	9.7	8.0	0.8	10.5	
					8.9	
P-LCR	[%]	18.4	60.0	11.0	29.4	
					7.4	
PCT	[%]	0.23	24.0	0.06	0.29	
					0.17	

1) These parameters values should be used as a reference only.