

SimplexTAS® 101 Reagent Cartridge

Control information

GLU

1. Reagent Cartridge information

Subject	Contents
Item	Glucose
Lot. No.	00209
Expired date	2021. 7. 20

2. Control information

Subject	Contents
Item	Lyphochek® Assayed Chemistry Control
REF No.	C-310-5 Level 1 C-315-5 Level 2
Lot. No.	26460
Manufacturer	Bio-rad

3. Control range

Units	Level 1 - 26461		Level 2 - 26462	
	Mean	Range	Mean	Range
mg/dL	77	66 ~ 89	253	215 ~ 291
mmol/L	4.3	3.7 ~ 4.9	14.0	11.9 ~ 16.2



TASCOM Co., Ltd

Anyang Megavalley, 268, Hagui-ro, Dongan-gu, Anyang-si, Gyeonggi-do, 14056, Korea

Email: info@tascom.org

Website: www.simplextas.com