

SimplexTAS® 101 Reagent Cartridge

Control information

TC

1. Reagent Cartridge information

Subject	Contents
Item	Total cholesterol
Lot. No.	00F09
Expired date	2021. 7. 8

2. Control information

Subject	Contents
Item	Lyphochek® Assayed Chemistry Control
REF No.	C-310-5 Level 1 C-315-5 Level 2
Lot. No.	26460
Manufacturer	Bio-rad

3. Control range

Units	Level 1 - 26461		Level 2 - 26462	
	Mean	Range	Mean	Range
mg/dL	268	227 ~ 308	113	96 ~ 130
mmol/L	6.9	5.9 ~ 8.0	2.9	2.5 ~ 3.4



TASCOM Co., Ltd

Anyang Megavalley, 268, Hagui-ro, Dongan-gu, Anyang-si, Gyeonggi-do, 14056, Korea

Email: info@tascom.org

Website: www.simplextas.com