

SimplexTAS® 101 Reagent Cartridge

Control information

CRP

1. Reagent Cartridge information

Subject	Contents
Item	CRP
Lot. No.	0160E
Expired date	2021. 8. 19

2. Control information

Subject	Contents
Item	Liquichek® Elevated CRP Control
REF No.	255 Level 2 256 Level 3
Lot. No.	35480
Manufacturer	Bio-rad
Expired date	2022. 12. 31

3. Control range

Units	Level 2 - 35482		Level 3 - 35483	
	Mean	Range	Mean	Range
mg/L	99.0	79.2 ~ 118.8	162.8	130.3 ~ 195.4
nmol/L	943	754 ~ 1131	1551	1241 ~ 1861



TASCOM Co., Ltd

Anyang Megavalley, 268, Hagui-ro, Dongan-gu, Anyang-si, Gyeonggi-do, 14056, Korea

Email: info@tascom.org

Website: www.simplextas.com