


















EIGHTCHECK-3WP ASSAY SHEET














High Level

For XP series

LOT 4111 0823 

 27-Jul-2024 

Parameters		Mean	Limit %	Limit #	Range	Barcode
WBC	[10 ³ /μL]	17.9	7.0	1.3	19.2	
					16.6	
RBC	[10 ⁶ /μL]	5.35	4.0	0.21	5.56	
					5.14	
HGB	[g/dL]	16.2	4.0	0.6	16.8	
					15.6	
	[mmol/L]	10.1	4.0	0.4	10.5	
					9.7	
HCT	[%]	45.0	7.5	3.4	48.4	
					41.6	
	[L/L]	0.450	7.5	0.034	0.484	
					0.416	
MCV	[fL]	84.2	5.0	4.2	88.4	
					80.0	
MCH	[pg]	30.3	5.5	1.7	32.0	
					28.6	
	[amol]	1879	5.5	103	1982	
					1776	
MCHC	[g/dL]	36.0	7.0	2.5	38.5	
					33.5	
	[mmol/L]	22.3	7.0	1.6	23.9	
					20.7	
PLT	[10 ³ /μL]	547	13	71	618	
					476	
W-SCR / LYM	[%]	35.5	15.0	5.3	40.8	
					30.2	

Parameters		Mean	Limit %	Limit #	Range	Barcode
W-MCR / MXD	[%]	16.1	25.0	4.0	20.1	
					12.1	
W-LCR / NEUT ¹⁾	[%]	48.4	7.0	3.4	51.8	
					45.0	
W-SCC / LYM ¹⁾	[10 ³ /μL]	6.4	15.0	1.0	7.4	
					5.4	
W-MCC / MXD ¹⁾	[10 ³ /μL]	2.9	25.0	0.7	3.6	
					2.2	
W-LCC / NEUT ¹⁾	[10 ³ /μL]	8.6	10.0	0.9	9.5	
					7.7	
W-SMV	[fL]	59.3	18.0	10.7	70.0	
					48.6	
W-LMV	[fL]	190.4	15.0	28.6	219.0	
					161.8	
RDW-SD	[fL]	35.7	15.0	5.4	41.1	
					30.3	
RDW-CV	[%]	10.2	15.0	1.5	11.7	
					8.7	
PDW	[fL]	8.5	10.0	0.9	9.4	
					7.6	
MPV	[fL]	9.3	8.0	0.7	10.0	
					8.6	
P-LCR	[%]	14.8	60.0	8.9	23.7	
					5.9	
PCT	[%]	0.51	20.0	0.10	0.61	
					0.41	

1) These parameters values should be used as a reference only.