

SimplexTAS® 101 Reagent Cartridge

Control information

HbA1c

1. Reagent Cartridge information

Subject	Contents
Item	HbA1c
Lot. No.	0010C
Expired date	2021. 3. 19

2. Control information

Subject	Contents
Item	Lyphochek® Diabetes Control
REF No.	740
Lot. No.	33980
Manufacturer	Bio-rad

3. Control range

Units	Level 1 - 33981		Level 2 - 33982	
	Mean	Range	Mean	Range
%	5.3	4.5 ~ 6.1	9.8	8.3 ~ 11.2
mmol/mol	34	26 ~ 43	84	67 ~ 99