
















EIGHTCHECK-3WP ASSAY SHEET














Low Level

For XP series

LOT 4111 0821 

Parameters		Mean	Limit %	Limit #	Range	Barcode
WBC	[10 ³ /μL]	3.2	12.0	0.4	3.6	
					2.8	
RBC	[10 ⁶ /μL]	2.67	5.0	0.13	2.80	
					2.54	
HGB	[g/dL]	6.1	5.0	0.3	6.4	
					5.8	
	[mmol/L]	3.8	5.0	0.2	4.0	
					3.6	
HCT	[%]	18.2	8.5	1.5	19.7	
					16.7	
	[L/L]	0.182	8.5	0.015	0.197	
					0.167	
MCV	[fL]	68.1	6.0	4.1	72.2	
					64.0	
MCH	[pg]	22.8	6.0	1.4	24.2	
					21.4	
	[amol]	1418	6.0	85	1503	
					1333	
MCHC	[g/dL]	33.5	8.0	2.7	36.2	
					30.8	
	[mmol/L]	20.8	8.0	1.7	22.5	
					19.1	
PLT	[10 ³ /μL]	82	30	25	107	
					57	
W-SCR / LYM	[%]	19.9	25.0	5.0	24.9	
					14.9	

 27-Jul-2024 

Parameters		Mean	Limit %	Limit #	Range	Barcode
W-MCR / MXD	[%]	10.0	30.0	3.0	13.0	
					7.0	
W-LCR / NEUT ¹⁾	[%]	70.1	15.0	10.5	80.6	
					59.6	
W-SCC / LYM ¹⁾	[10 ³ /μL]	0.6	25.0	0.2	0.8	
					0.4	
W-MCC / MXD ¹⁾	[10 ³ /μL]	0.3	30.0	0.1	0.4	
					0.2	
W-LCC / NEUT ¹⁾	[10 ³ /μL]	2.3	20.0	0.5	2.8	
					1.8	
W-SMV	[fL]	58.9	20.0	11.8	70.7	
					47.1	
W-LMV	[fL]	189.3	20.0	37.9	227.2	
					151.4	
RDW-SD	[fL]	40.9	20.0	8.2	49.1	
					32.7	
RDW-CV	[%]	15.7	20.0	3.1	18.8	
					12.6	
PDW	[fL]	8.4	20.0	1.7	10.1	
					6.7	
MPV	[fL]	9.2	10.0	0.9	10.1	
					8.3	
P-LCR	[%]	15.8	60.0	9.5	25.3	
					6.3	
PCT	[%]	0.08	43.0	0.03	0.11	
					0.05	

1) These parameters values should be used as a reference only.