
















EIGHTCHECK-3WP ASSAY SHEET













Normal Level

For pocH-100i

LOT 4111 0822 

Parameters		Mean	Limit %	Limit #	Range	Barcode
WBC	[10 ³ /μL]	6.9	9.0	0.6	7.5	
					6.3	
RBC	[10 ⁶ /μL]	4.59	5.0	0.23	4.82	
					4.36	
HGB	[g/dL]	12.2	5.0	0.6	12.8	
					11.6	
	[mmol/L]	7.6	5.0	0.4	8.0	
					7.2	
HCT	[%]	36.6	8.0	2.9	39.5	
					33.7	
	[L/L]	0.366	8.0	0.029	0.395	
0.337						
MCV	[fL]	79.8	6.0	4.8	84.6	
					75.0	
MCH	[pg]	26.6	6.0	1.6	28.2	
					25.0	
	[amol]	1650	6.0	99	1749	
1551						
MCHC	[g/dL]	33.3	7.0	2.3	35.6	
					31.0	
	[mmol/L]	20.7	7.0	1.4	22.1	
					19.3	
PLT	[10 ³ /μL]	216	15	32	248	
					184	
W-SCR / LYM	[%]	32.1	15.0	4.8	36.9	
					27.3	

 27-Jul-2024 

Parameters		Mean	Limit %	Limit #	Range	Barcode
W-MCR / MXD	[%]	11.8	28.0	3.3	15.1	
					8.5	
W-LCR / NEUT ¹⁾	[%]	56.1	7.0	3.9	60.0	
					52.2	
W-SCC / LYM ¹⁾	[10 ³ /μL]	2.2	15.0	0.3	2.5	
					1.9	
W-MCC / MXD ¹⁾	[10 ³ /μL]	0.8	28.0	0.2	1.0	
					0.6	
W-LCC / NEUT ¹⁾	[10 ³ /μL]	3.9	12.0	0.5	4.4	
					3.4	
W-SMV	[fL]	63.9	18.0	11.5	75.4	
					52.4	
W-LMV	[fL]	200.5	15.0	30.1	230.6	
					170.4	
RDW-SD	[fL]	44.6	25.0	11.2	55.8	
					33.4	
RDW-CV	[%]	16.2	25.0	4.1	20.3	
					12.1	
PDW	[fL]	9.1	10.0	0.9	10.0	
					8.2	
MPV	[fL]	10.1	8.0	0.8	10.9	
					9.3	
P-LCR	[%]	17.3	60.0	10.4	27.7	
					6.9	

1) These parameters values should be used as a reference only.

