
















EIGHTCHECK-3WP ASSAY SHEET













Low Level

For pocH-100i

LOT 4111 0821 

Parameters		Mean	Limit %	Limit #	Range	Barcode
WBC	[10 ³ /μL]	3.1	14.0	0.4	3.5	
					2.7	
RBC	[10 ⁶ /μL]	2.71	6.0	0.16	2.87	
					2.55	
HGB	[g/dL]	6.2	6.0	0.4	6.6	
					5.8	
	[mmol/L]	3.8	6.0	0.2	4.0	
					3.6	
HCT	[%]	19.1	8.5	1.6	20.7	
					17.5	
	[L/L]	0.191	8.5	0.016	0.207	
					0.175	
MCV	[fL]	70.6	7.0	4.9	75.5	
					65.7	
MCH	[pg]	22.9	7.0	1.6	24.5	
					21.3	
	[amol]	1420	7.0	99	1519	
					1321	
MCHC	[g/dL]	32.5	8.0	2.6	35.1	
					29.9	
	[mmol/L]	20.1	8.0	1.6	21.7	
					18.5	
PLT	[10 ³ /μL]	78	30	23	101	
					55	
W-SCR / LYM	[%]	20.3	25.0	5.1	25.4	
					15.2	

 27-Jul-2024 

Parameters		Mean	Limit %	Limit #	Range	Barcode
W-MCR / MXD	[%]	10.0	48.0	4.8	14.8	
					5.2	
W-LCR / NEUT ¹⁾	[%]	69.7	15.0	10.5	80.2	
					59.2	
W-SCC / LYM ¹⁾	[10 ³ /μL]	0.6	30.0	0.2	0.8	
					0.4	
W-MCC / MXD ¹⁾	[10 ³ /μL]	0.3	41.0	0.1	0.4	
					0.2	
W-LCC / NEUT ¹⁾	[10 ³ /μL]	2.2	20.0	0.4	2.6	
					1.8	
W-SMV	[fL]	64.0	20.0	12.8	76.8	
					51.2	
W-LMV	[fL]	200.1	20.0	40.0	240.1	
					160.1	
RDW-SD	[fL]	45.9	30.0	13.8	59.7	
					32.1	
RDW-CV	[%]	18.9	30.0	5.7	24.6	
					13.2	
PDW	[fL]	8.9	20.0	1.8	10.7	
					7.1	
MPV	[fL]	10.4	10.0	1.0	11.4	
					9.4	
P-LCR	[%]	19.8	60.0	11.9	31.7	
					7.9	

1) These parameters values should be used as a reference only.

