
















EIGHTCHECK-3WP ASSAY SHEET













Normal Level

For pocH-100i

LOT 4027 0822 

Parameters		Mean	Limit %	Limit #	Range	Barcode
WBC	[10 ³ /μL]	7.3	9.0	0.7	8.0	
					6.6	
RBC	[10 ⁶ /μL]	4.56	5.0	0.23	4.79	
					4.33	
HGB	[g/dL]	12.6	5.0	0.6	13.2	
					12.0	
	[mmol/L]	7.8	5.0	0.4	8.2	
					7.4	
HCT	[%]	37.5	8.0	3.0	40.5	
					34.5	
	[L/L]	0.375	8.0	0.030	0.405	
					0.345	
MCV	[fL]	82.3	6.0	4.9	87.2	
					77.4	
MCH	[pg]	27.6	6.0	1.7	29.3	
					25.9	
	[amol]	1715	6.0	103	1818	
					1612	
MCHC	[g/dL]	33.6	7.0	2.4	36.0	
					31.2	
	[mmol/L]	20.9	7.0	1.5	22.4	
					19.4	
PLT	[10 ³ /μL]	223	15	33	256	
					190	
W-SCR / LYM	[%]	30.6	15.0	4.6	35.2	
					26.0	

 **04-May-2024** 

Parameters		Mean	Limit %	Limit #	Range	Barcode
W-MCR / MXD	[%]	12.5	28.0	3.5	16.0	
					9.0	
W-LCR / NEUT ¹⁾	[%]	56.9	7.0	4.0	60.9	
					52.9	
W-SCC / LYM ¹⁾	[10 ³ /μL]	2.2	15.0	0.3	2.5	
					1.9	
W-MCC / MXD ¹⁾	[10 ³ /μL]	0.9	28.0	0.3	1.2	
					0.6	
W-LCC / NEUT ¹⁾	[10 ³ /μL]	4.2	12.0	0.5	4.7	
					3.7	
W-SMV	[fL]	64.6	18.0	11.6	76.2	
					53.0	
W-LMV	[fL]	204.2	15.0	30.6	234.8	
					173.6	
RDW-SD	[fL]	42.5	25.0	10.6	53.1	
					31.9	
RDW-CV	[%]	15.0	25.0	3.8	18.8	
					11.2	
PDW	[fL]	9.7	10.0	1.0	10.7	
					8.7	
MPV	[fL]	10.5	8.0	0.8	11.3	
					9.7	
P-LCR	[%]	21.8	60.0	13.1	34.9	
					8.7	

1) These parameters values should be used as a reference only.

