

# SimplexTAS® 101 Reagent Cartridge

## Control information

### CRE

#### 1. Reagent Cartridge information

Subject	Contents
Item	Creatinine
Lot. No.	0050F
Expired date	2021. 3. 31

#### 2. Control information

Subject	Contents
Item	Lyphochek® Assayed Chemistry Control
REF No.	C-310-5 Level 1 C-315-5 Level 2
Lot. No.	26460
Manufacturer	Bio-rad

#### 3. Control range

Units	Level 1 - 26461		Level 2 - 26462	
	Mean	Range	Mean	Range
mg/dL	2.7	2.3 ~ 3.1	5.7	4.8 ~ 6.5
umol/L	239	203 ~ 274	504	424 ~ 575



TASCOM Co., Ltd

Anyang Megavalley, 268, Hagui-ro, Dongan-gu, Anyang-si, Gyeonggi-do, 14056, Korea

**Email:** info@tascom.org

**Website:** www.simplextas.com