

# SimplexTAS® 101 Reagent Cartridge

## Control information

### GGT(P)

#### 1. Reagent Cartridge information

Subject	Contents
Item	GGT (Plasma sample)
Lot. No.	00B84
Expired date	2021. 7. 21

#### 2. Control information

Subject	Contents
Item	Lyphochek® Assayed Chemistry Control
REF No.	C-310-5 Level 1 C-315-5 Level 2
Lot. No.	26460
Manufacturer	Bio-rad

#### 3. Control range

Units	Level 1 - 26461		Level 2 - 26462	
	Mean	Range	Mean	Range
IU/L	61	52 ~ 70	168	142 ~ 193
ukat/L	1.02	0,87 ~ 1.17	2.81	2.37 ~ 3.22



TASCOM Co., Ltd

Anyang Megavalley, 268, Hagui-ro, Dongan-gu, Anyang-si, Gyeonggi-do, 14056, Korea

**Email:** info@tascom.org

**Website:** www.simplextas.com