

# EIGHTCHECK-3WP ASSAY SHEET

## High Level

### For poCH series



**LOT** 4279 0823

**11-Jan-2025**

Parameters	Mean	Limit %	Limit #	Range
WBC [10 <sup>3</sup> /μL]	17.3	9.0	1.6	18.9 15.7

Parameters	Mean	Limit %	Limit #	Range
PDW [fL]	9.6	10.0	1.0	10.6 8.6

RBC [10 <sup>6</sup> /μL]	5.51	5.0	0.28	5.79 5.23

MPV [fL]	10.7	8.0	0.9	11.6 9.8

MCV [fL]	86.5	6.0	5.2	91.7 81.3

P-LCR [%]	21.9	60.0	13.1	35.0 8.8

PLT [10 <sup>3</sup> /μL]	499	15.0	75	574 424

W-SCR / LYM [%]	35.6	15.0	5.3	40.9 30.3

W-MCR / MXD [%]	14.3	28.0	4.0	18.3 10.3

W-LCR / NEUT <sup>1)</sup> [%]	50.1	7.0	3.5	53.6 46.6

W-SCC / LYM <sup>1)</sup> [10 <sup>3</sup> /μL]	6.2	15.0	0.9	7.1 5.3

W-MCC / MXD <sup>1)</sup> [10 <sup>3</sup> /μL]	2.5	28.0	0.7	3.2 1.8

W-LCC / NEUT <sup>1)</sup> [10 <sup>3</sup> /μL]	8.6	12.0	1.0	9.6 7.6

W-SMV [fL]	63.0	18.0	11.3	74.3 51.7

W-LMV [fL]	188.1	15.0	28.2	216.3 159.9

RDW-SD [fL]	42.4	25.0	10.6	53.0 31.8

RDW-CV [%]	14.1	25.0	3.5	17.6 10.6

Conventional Units				
Parameters	Mean	Limit %	Limit #	Range
HGB [g/dL]	16.3	5.0	0.8	17.1 15.5

HCT [%]	47.7	8.0	3.8	51.5 43.9

MCH [pg]	29.6	5.5	1.6	31.2 28.0

MCHC [g/dL]	34.2	7.0	2.4	36.6 31.8

SI Units				
Parameters	Mean	Limit %	Limit #	Range
HGB [mmol/L]	10.1	5.0	0.5	10.6 9.6

HCT [L/L]	0.477	8.0	0.038	0.515 0.439

MCH [amol]	1836	5.5	101	1937 1735

MCHC [mmol/L]	21.2	7.0	1.5	22.7 19.7

<sup>1)</sup> These parameters values should be used as a reference only.

Revised 08/2024

