

SimplexTAS® 101 Reagent Cartridge

Control information

Hb

1. Reagent Cartridge information

Subject	Contents
Item	Hemoglobin
Lot. No.	02A05
Expired date	2021. 7. 20

2. Control information

Subject	Contents
Item	Meter Trax™ Control
REF No.	970
Lot. No.	92790
Manufacturer	Bio-rad

3. Control range

Units	Level 1 - 92791		Level 2 - 92793	
	Mean	Range	Mean	Range
g/dL	7.4	6.3 ~ 8.6	15.1	12.8 ~ 17.3
mmol/L	4.59	3.91 ~ 5.34	9.37	7.94 ~ 10.74



TASCOM Co., Ltd

Anyang Megavalley, 268, Hagui-ro, Dongan-gu, Anyang-si, Gyeonggi-do, 14056, Korea

Email: info@tascom.org

Website: www.simplextas.com