

# SimplexTAS<sup>®</sup> 101 Reagent Cartridge

## Control information

### BUN

#### 1. Reagent Cartridge information

Subject	Contents
Item	BUN
Lot. No.	0080C
Expired date	2021. 3. 30

#### 2. Control information

Subject	Contents
Item	Lyphochek <sup>®</sup> Assayed Chemistry Control
REF No.	C-310-5 Level 1 C-315-5 Level 2
Lot. No.	26460
Manufacturer	Bio-rad

#### 3. Control range

Units	Level 1 - 26461		Level 2 - 26462	
	Mean	Range	Mean	Range
mg/dL	17	14 ~ 20	49	41 ~ 56
mmol/L	6.1	5.0 ~ 7.1	17.5	14.6 ~ 20.0



TASCOM Co., Ltd

Anyang Megavalley, 268, Hagui-ro, Dongan-gu, Anyang-si, Gyeonggi-do, 14056, Korea

**Email:** info@tascom.org

**Website:** www.simplextas.com