

SimplexTAS® 101 Reagent Cartridge

Control information

ALT

1. Reagent Cartridge information

Subject	Contents
Item	ALT(GPT)
Lot. No.	0090B
Expired date	2021. 4. 20

2. Control information

Subject	Contents
Item	Lyphochek® Assayed Chemistry Control
REF No.	C-310-5 Level 1 C-315-5 Level 2
Lot. No.	26460
Manufacturer	Bio-rad

3. Control range

Units	Level 1 - 26461		Level 2 - 26462	
	Mean	Range	Mean	Range
IU/L	23	20 ~ 27	97	83 ~ 112
ukat/L	0.38	0.33 ~ 0.45	1.62	1.39 ~ 1.87



TASCOM Co., Ltd

Anyang Megavalley, 268, Hagui-ro, Dongan-gu, Anyang-si, Gyeonggi-do, 14056, Korea

Email: info@tascom.org

Website: www.simplextas.com