

SimplexTAS[®] 101 Reagent Cartridge

Control information

HDL

1. Reagent Cartridge information

Subject	Contents
Item	HDL
Lot. No.	0110A
Expired date	2021. 7. 7

2. Control information

Subject	Contents
Item	Lyphochek [®] Assayed Chemistry Control
REF No.	C-310-5 Level 1 C-315-5 Level 2
Lot. No.	26460
Manufacturer	Bio-rad

3. Control range

Units	Level 1 - 26461		Level 2 - 26462	
	Mean	Range	Mean	Range
mg/dL	62	53 ~ 72	32	27 ~ 37
mmol/L	1.60	1.37 ~ 1.86	0.83	0.70 ~ 0.96



TASCOM Co., Ltd

Anyang Megavalley, 268, Hagui-ro, Dongan-gu, Anyang-si, Gyeonggi-do, 14056, Korea

Email: info@tascom.org

Website: www.simplextas.com