

CB871331D  
Part No 350608

# EIGHTCHECK-3WP ASSAY SHEET

## Low Level For XP series



**LOT** 3140 0821



**26-Aug-2023**



Parameters		Mean	Limit %	Limit #	Range	Barcode
WBC	[10 <sup>3</sup> /μL]	3.4	12.0	0.4	3.8	
					3.0	
RBC	[10 <sup>6</sup> /μL]	2.63	5.0	0.13	2.76	
					2.50	
HGB	[g/dL]	6.8	5.0	0.3	7.1	
					6.5	
	[mmol/L]	4.2	5.0	0.2	4.4	
					4.0	
HCT	[%]	19.8	8.5	1.7	21.5	
					18.1	
	[L/L]	0.198	8.5	0.017	0.215	
					0.181	
MCV	[fL]	75.2	6.0	4.5	79.7	
					70.7	
MCH	[pg]	25.9	6.0	1.6	27.5	
					24.3	
	[amol]	1605	6.0	96	1701	
					1509	
MCHC	[g/dL]	34.3	8.0	2.7	37.0	
					31.6	
	[mmol/L]	21.3	8.0	1.7	23.0	
					19.6	
PLT	[10 <sup>3</sup> /μL]	85	30	26	111	
					59	
W-SCR / LYM	[%]	19.6	25.0	4.9	24.5	
					14.7	

Parameters		Mean	Limit %	Limit #	Range	Barcode
W-MCR / MXD	[%]	10.4	30.0	3.1	13.5	
					7.3	
W-LCR / NEUT <sup>1)</sup>	[%]	70.0	15.0	10.5	80.5	
					59.5	
W-SCC / LYM <sup>1)</sup>	[10 <sup>3</sup> /μL]	0.7	25.0	0.2	0.9	
					0.5	
W-MCC / MXD <sup>1)</sup>	[10 <sup>3</sup> /μL]	0.4	30.0	0.1	0.5	
					0.3	
W-LCC / NEUT <sup>1)</sup>	[10 <sup>3</sup> /μL]	2.3	20.0	0.5	2.8	
					1.8	
W-SMV	[fL]	57.7	20.0	11.5	69.2	
					46.2	
W-LMV	[fL]	192.6	20.0	38.5	231.1	
					154.1	
RDW-SD	[fL]	37.7	20.0	7.5	45.2	
					30.2	
RDW-CV	[%]	12.4	20.0	2.5	14.9	
					9.9	
PDW	[fL]	8.6	20.0	1.7	10.3	
					6.9	
MPV	[fL]	9.4	10.0	0.9	10.3	
					8.5	
P-LCR	[%]	16.4	60.0	9.8	26.2	
					6.6	
PCT	[%]	0.08	43.0	0.03	0.11	
					0.05	

1) These parameters values should be used as a reference only.