














AH951421D  
Part No 350462

# EIGHTCHECK-3WP ASSAY SHEET













## Low Level For pochH-100i



**LOT** 3140 0821 

Parameters	Mean	Limit %	Limit #	Range	Barcode
WBC [10 <sup>3</sup> /μL]	3.3	14.0	0.5	3.8	
				2.8	
RBC [10 <sup>6</sup> /μL]	2.70	6.0	0.16	2.86	
				2.54	
HGB [g/dL] [mmol/L]	6.8	6.0	0.4	7.2	
				6.4	
	4.2	6.0	0.3	4.5	
				3.9	
HCT [%] [L/L]	21.1	8.5	1.8	22.9	
				19.3	
	0.211	8.5	0.018	0.229	
				0.193	
MCV [fL]	78.2	7.0	5.5	83.7	
				72.7	
MCH [pg] [amol]	25.2	7.0	1.8	27.0	
				23.4	
	1563	7.0	109	1672	
				1454	
MCHC [g/dL] [mmol/L]	32.2	8.0	2.6	34.8	
				29.6	
	20.0	8.0	1.6	21.6	
				18.4	
PLT [10 <sup>3</sup> /μL]	80	30	24	104	
				56	
W-SCR / LYM	[%]	19.2	25.0	4.8	
				24.0	
				14.4	

 26-Aug-2023 

Parameters	Mean	Limit %	Limit #	Range	Barcode
W-MCR / MXD [%]	9.3	48.0	4.5	13.8	
				4.8	
W-LCR / NEUT <sup>1)</sup> [%]	71.5	15.0	10.7	82.2	
				60.8	
W-SCC / LYM <sup>1)</sup> [10 <sup>3</sup> /μL]	0.6	30.0	0.2	0.8	
				0.4	
W-MCC / MXD <sup>1)</sup> [10 <sup>3</sup> /μL]	0.3	41.0	0.1	0.4	
				0.2	
W-LCC / NEUT <sup>1)</sup> [10 <sup>3</sup> /μL]	2.4	20.0	0.5	2.9	
				1.9	
W-SMV [fL]	61.7	20.0	12.3	74.0	
				49.4	
W-LMV [fL]	200.9	20.0	40.2	241.1	
				160.7	
RDW-SD [fL]	43.7	30.0	13.1	56.8	
				30.6	
RDW-CV [%]	16.2	30.0	4.9	21.1	
				11.3	
PDW [fL]	8.9	20.0	1.8	10.7	
				7.1	
MPV [fL]	10.5	10.0	1.1	11.6	
				9.4	
P-LCR [%]	19.8	60.0	11.9	31.7	
				7.9	

1) These parameters values should be used as a reference only.

