

CD654074D
Part No 350610

EIGHTCHECK-3WP ASSAY SHEET

High Level For XP series



LOT 2085 0823



Parameters		Mean	Limit %	Limit #	Range	Barcode
WBC	[10 ³ /μL]	17.9	7.0	1.3	19.2	
					16.6	
RBC	[10 ⁶ /μL]	5.29	4.0	0.21	5.50	
					5.08	
HGB	[g/dL]	16.7	3.5	0.6	17.3	
					16.1	
	[mmol/L]	10.4	3.5	0.4	10.8	
					10.0	
HCT	[%]	46.6	7.5	3.5	50.1	
					43.1	
	[L/L]	0.466	7.5	0.035	0.501	
					0.431	
MCV	[fL]	88.0	5.0	4.4	92.4	
					83.6	
MCH	[pg]	31.6	5.5	1.7	33.3	
					29.9	
	[amol]	1959	5.5	108	2067	
					1851	
MCHC	[g/dL]	35.8	7.0	2.5	38.3	
					33.3	
	[mmol/L]	22.2	7.0	1.6	23.8	
					20.6	
PLT	[10 ³ /μL]	559	13	73	632	
					486	
W-SCR / LYM	[%]	35.1	15.0	5.3	40.4	
					29.8	

02-Jul-2022



Parameters		Mean	Limit %	Limit #	Range	Barcode
W-MCR / MXD	[%]	16.5	25.0	4.1	20.6	
					12.4	
W-LCR / NEUT ¹⁾	[%]	48.4	7.0	3.4	51.8	
					45.0	
W-SCC / LYM ¹⁾	[10 ³ /μL]	6.3	15.0	0.9	7.2	
					5.4	
W-MCC / MXD ¹⁾	[10 ³ /μL]	3.0	25.0	0.8	3.8	
					2.2	
W-LCC / NEUT ¹⁾	[10 ³ /μL]	8.6	10.0	0.9	9.5	
					7.7	
W-SMV	[fL]	59.9	18.0	10.8	70.7	
					49.1	
W-LMV	[fL]	195.7	15.0	29.4	225.1	
					166.3	
RDW-SD	[fL]	33.4	15.0	5.0	38.4	
					28.4	
RDW-CV	[%]	8.9	15.0	1.3	10.2	
					7.6	
PDW	[fL]	8.8	10.0	0.9	9.7	
					7.9	
MPV	[fL]	9.8	8.0	0.8	10.6	
					9.0	
P-LCR	[%]	18.1	60.0	10.9	29.0	
					7.2	
PCT	[%]	0.55	20.0	0.11	0.66	
					0.44	

1) These parameters values should be used as a reference only.