

CD654074D
Part No 350610

EIGHTCHECK-3WP ASSAY SHEET

High Level For XP series



LOT 3140 0823



26-Aug-2023



Parameters	Mean	Limit %	Limit #	Range	Barcode
WBC [10 ³ /μL]	17.6	7.0	1.2	18.8	
				16.4	
RBC [10 ⁶ /μL]	5.29	4.0	0.21	5.50	
				5.08	
HGB [g/dL] [mmol/L]	16.9	4.0	0.7	17.6	
				16.2	
	10.5	4.0	0.4	10.9	
				10.1	
HCT [%] [L/L]	47.4	7.5	3.6	51.0	
				43.8	
	0.474	7.5	0.036	0.510	
			0.438		
MCV [fL]	89.6	5.0	4.5	94.1	
				85.1	
MCH [pg] [amol]	31.9	5.5	1.8	33.7	
				30.1	
	1983	5.5	109	2092	
				1874	
MCHC [g/dL] [mmol/L]	35.7	7.0	2.5	38.2	
				33.2	
	22.1	7.0	1.5	23.6	
				20.6	
PLT [10 ³ /μL]	565	13	73	638	
				492	
W-SCR / LYM [%]	35.5	15.0	5.3	40.8	
				30.2	

Parameters	Mean	Limit %	Limit #	Range	Barcode
W-MCR / MXD [%]	16.2	25.0	4.1	20.3	
				12.1	
W-LCR / NEUT ¹⁾ [%]	48.3	7.0	3.4	51.7	
				44.9	
W-SCC / LYM ¹⁾ [10 ³ /μL]	6.2	15.0	0.9	7.1	
				5.3	
W-MCC / MXD ¹⁾ [10 ³ /μL]	2.9	25.0	0.7	3.6	
				2.2	
W-LCC / NEUT ¹⁾ [10 ³ /μL]	8.5	10.0	0.9	9.4	
				7.6	
W-SMV [fL]	58.4	18.0	10.5	68.9	
				47.9	
W-LMV [fL]	193.5	15.0	29.0	222.5	
				164.5	
RDW-SD [fL]	37.2	15.0	5.6	42.8	
				31.6	
RDW-CV [%]	9.8	15.0	1.5	11.3	
				8.3	
PDW [fL]	8.7	10.0	0.9	9.6	
				7.8	
MPV [fL]	9.4	8.0	0.8	10.2	
				8.6	
P-LCR [%]	16.0	60.0	9.6	25.6	
				6.4	
PCT [%]	0.53	20.0	0.11	0.64	
				0.42	

1) These parameters values should be used as a reference only.